

EU-Konformitätserklärung

entsprechend EN ISO/IEC 17050-1

PRÜFTECHNIK Dieter Busch GmbH,
Freisinger Str. 34, 85737 Ismaning, Deutschland
erklärt, dass das Produkt

Name: **Batteriefach**
Modellnummer: **ALI 4.605**
Typ: **Ausrichtsystem**

mit den zutreffenden Europäischen Richtlinien konform ist. Die in den Richtlinien festgelegten wesentlichen Schutzanforderungen werden eingehalten.


Richtlinien

2014/30/EU	Elektromagnetische Verträglichkeit
2011/65/EU	Richtlinie zur Beschränkung der Verwendung bestimmter gefährlicher Stoffe in Elektro- und Elektronikgeräten (inkl. Ergänzungen aus 2015/863/EU)

Angewandte Normen

EN 61326-1:2013	Elektrische Mess-, Steuer-, Regel- und Laborgeräte - EMC- Anforderungen - Teil 1: Allgemeine Anforderungen
EN 61326-2-2:2013	El. Mess-, Steuer-, Regel- und Laborgeräte - EMC- Anf. - Teil 2-2: Besondere Anf. - Prüfanordnung, Betriebsbedingungen und Leistungsmerkmale für ortsveränderliche Prüf-, Mess- und Überwachungsgeräte für Gebrauch in Niederspannungs- Stromversorgungsnetzen

Ismaning, 14. Dezember 2021
Ort, Datum der Ausstellung


Dr. Edwin Becker, Geschäftsführer

DOC 46.501



EU Declaration of Conformity

in accordance with EN ISO/IEC 17050-1

PRÜFTECHNIK Dieter Busch GmbH,
 Freisinger Str. 34, 85737 Ismaning, Germany
 declares, that the equipment

Name: **Disposable Battery Housing**

Model Number: **ALI 4.605**

Type: **Alignment System**

complies with the appropriate European Directives. The essential safety requirements set out in the European Directives are fulfilled.

Directives


2014/30/EU	Electromagnetic compatibility
2011/65/EU	Directive on the restriction of the use of certain hazardous substances in electrical and electronic equipment (incl. amendments from 2015/863/EU)

Standards applied

EN 61326-1:2013	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements
EN 61326-2-2:2013	El. equip for measurement, control and lab. use - EMC -Part 2-2: Particular requirements - Test config., operational conditions and performance criteria for portable test, measuring and monitoring equip used in low-voltage distribution systems

Ismaning, 14. Dezember 2021

Place, date of issue



Dr. Edwin Becker, Managing Director

DOC 46.501

