

# EU-Konformitätserklärung

entsprechend EN ISO/IEC 17050-1

**PRÜFTECHNIK Dieter Busch AG,**  
Oskar-Messter Str. 19-21, 85737 Ismaning, Deutschland  
erklärt, dass das Produkt

Name: **Keyphaser adapter**  
Modellnummer: **VIB 5.332 X**  
Typ: **Condition Monitoring System**

Gültig ab:

mit den zutreffenden Europäischen Richtlinien konform ist. Die in den Richtlinien festgelegten wesentlichen Schutzanforderungen werden eingehalten.

## Richtlinien

2014/30/EU	Elektromagnetische Verträglichkeit
2011/65/EU	Beschränkung der Verwendung bestimmter gefährlicher Stoffe in Elektro- und Elektronikgeräten (RoHS 2)

## Angewandte Normen

EN 61326-1:2013	Elektrische Mess-, Steuer-, Regel- und Laborgeräte - EMV- Anforderungen - Teil 1: Allgemeine Anforderungen
EN 61326-2-2:2013	El. Mess-, Steuer-, Regel- und Laborgeräte - EMV- Anf. - Teil 2-2: Besondere Anf. - Prüfanordnung, Betriebsbedingungen und Leistungsmerkmale für ortsveränderliche Prüf-, Mess- und Überwachungsgeräte für Gebrauch in Niederspannungs- Stromversorgungsnetzen

## Richtlinienspezifische Angaben



Ismaning, 24. November 2017

Ort, Datum der Ausstellung



Dr. Thomas Rohe - Vorstand

# EC Declaration of conformity

in accordance with EN ISO/IEC 17050-1

**PRÜFTECHNIK Dieter Busch AG,**  
Oskar-Messter Str. 19-21, 85737 Ismaning, Germany  
declares, that the equipment

Name: **Keyphaser adapter for protection system**

Model Number: **VIB 5.332 X**

Type: **Condition Monitoring System**

Valid from:

complies with the appropriate European Directives. The essential safety requirements set out in the European Directives are fulfilled.

#### Directives

<b>2014/30/EU</b>	Electromagnetic compatibility
<b>2011/65/EU</b>	Restriction of the use of certain hazardous substances in electrical and electronic equipment (RoHS 2)

#### Standards applied

<b>EN 61326-1:2013</b>	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements
<b>EN 61326-2-2:2013</b>	El. equip for measurement, control and lab. use - EMC -Part 2-2: Particular requirements - Test config., operational conditions and performance criteria for portable test, measuring and monitoring equip used in low-voltage distribution systems

#### Directive specific information



Ismaning, 24. November 2017

Place, date of issue



Dr. Thomas Rohe - Executive Board